

Phone 91-7232-242456

Fax 91-7232-244148

SHRI. VASANTRAO NAIK GOVT. MEDICAL COLLEGE, YAVATMAL.

श्री. वसंतराव नाईक शासकीय वैद्यकीय महाविद्यालय, यवतमाळ

Ref. No.SVNGMC/PLAN/19-20/QT/ ४६६ /20

Dt. १४/01/2020

Open Quotation Notice

For the supply of Equipments (SECOND EXTENTION)

Dean Shri Vasantrao Naik Govt. Medical College & Hospital, Yavatmal invited separate open quotation for the supply of Instruments **listed below** in **two envelope** system from manufacturer & authorized dealers. Detail list & specification is enclosed. Interested Manufacturer/dealer can send their quotation.

ENVELOPE NO :- 1 TECHNICAL BID:- Following documents must be submit in this envelope

1. Shop registration Certificate

In case of proprietorship / Partnership firms, names of proprietors / partners/ Directors with address and percentage of share Ownership status of the firm

(Maharashtra Govt. / Central Govt./Jt. Sector /co - operative /SSI /Private)

2. Copy of the Pan card

3. GST Registration Certificate

4. Catalogue/Leaflet of the Model

5. Compliance chart in the prescribe format

6. Authorization certificate from the manufacturer

7. User list of certificates of satisfactory performance.

8. Acceptance of all terms of conditions with stamp of signature firm/company on letter head.

9. Name & Phone no. who should be contacted by this office in case of any urgent problem.

10. Email Address Office contact number of the firm.

ENVELOPE NO :- 2 PRICE BID

1. Quote your Minimum Rates in a following Format on the letterhead of the firm/company in Price Bid envelop the cost should be quoted in Indian currency only. Quote maximum two models on separate sheet.

(i) Basic Price:-.....

(ii) Taxes: Excise duty..... Custom duty..... GST..... other.....

(iii) Installation Charges:-.....

(iv) TOTAL=(i)+(ii)+(iii)+(iv):-.....

(v) CMC/AMC:-.....

(vi) Prices for consumables/ Disposable/ accessories should be submitted separately in price bid which will be valid for warranty period, AMC.CMC period.

Both the sealed envelope No. 1 and No.2 should be put together in one cover and it should be sealed. This sealed envelope should be marked on upper left hand corner mentioning name of the items No of the Item Quotation enquiry No. And Date of quotation opening. The full name and address of the Firm/Company should be written on the left hand lower corner. Such separate envelope containing sets of Envelope No.1 and No. 2 should be submitted separately for each Items mentioned in the List.

Envelope No.2 will be opened only when the documents as mentioned in envelope No.1 are in order as per requirements and if item is as per specifications.

Separate quotation should be prepared for every item mentioned in the list and sent in sealed envelope (duly sealed) by post or hand delivery so as to reach Inward Section of administrative office, Shri Vasantrao Naik Govt. Medical College, Yavatmal on or before **Dt. 21 / 01 /2020** up to 05.30 pm.

Date & timing of the opening of quotation (Technical Bid) :- **Dt. 22/01/2020** at **11.00 am.** at College Council Hall. Date & time of opening of the Price bid will be decide after scrutiny of the technical detail. Representative of the Supplier may attend the meeting by submitting authorization letter.

Please quote your minimum rates for Supply of articles as per details given below/list enclosed,. The separate envelopes (every Item) containing the quotation should be sealed and super-scribed as under:-

Quotation due on. **14.01.20**. Against inquiry No. **486** for the purchase of Equipments.

LIST OF INSTRUMENT REQUIRED

Sr No	Name of Dept	Name of the Instrument/Equipment	Qtn Req.	Approx Amount
1	Biochemistry	Single door refrigerator	1	59000
2	Biochemistry	Medical clinical centrifuge brushless 48	1	45500

The terms and conditions for the supply are:-

1. Enquiry No. must invariably be given at the top of the envelope. Envelopes without indication will not be entertained.
2. For equipment/ accessories/ spares, technical bid and price bid must be sent in two separate, sealed envelopes and front of the envelop marked clearly in capitals as 'PRICE BID', 'TECHNICAL BID'. Otherwise the bid will be summarily REJECTED.
3. Offers other than those from the manufacturer should be supported by an authority letter from the manufacturer as per Annexure-A.
4. AMC/CMC/Extended Warranty if applicable must be quoted along with price bid and not with the Technical bid. However, in technical bid form consent for AMC/CMC must be given.CMC (Labor& Spare parts), these charges include the cost of replacement parts/Components/spares/accessories etc. This shall be borne by the firm. In case of consumables/these will also be included in CMC as per Technical Specification. However, AMC charge will cover labor only and does include the cost of any replacement parts/components etc.
5. **Unsealed quotations will be rejected.**
6. Full description, make and quantity of the articles must be mentioned in your quotation. Original catalogue/literature should be submitted with the quotation (Technical bid envelop)
7. Validity of the quotation should be for a minimum period of 120 days.
8. Delivery period for each item be specifically mentioned in the quotations from the date of Supply order it should be maximum of 6 weeks.
9. The Dean reserves the right to call the firm for physical demonstration of item to verify the quality of the offered item from the lowest bidder. The firm must demonstrate the product within 7 days of requisition, failing which the bid will be rejected.
10. The institute reserves the right to reject the supplied goods if the same are not found in accordance with the specifications approved and asked for. In case there is a short/defective supply the firm will be informed and the defective material will be lifted from the concerned department/Central store by the supplier at their own cost within two weeks period. The institute will not bear any expenses on this account. **The Hon'ble dean reserves the right to cancel the requested tender or quotation process at any time.**
11. **The rates and total cost must be quoted in both words and figures (over-writing not allowed.)**
12. Rate should be for free delivery at Shri V.N. Govt. Medical College, Yavatmal premises only. Rate should be quoted including all charges. Rates must be quote on letter head alongwith Pan, VAT, CST/GST Number. It should be submit in Price envelope. (envelope No 2)
13. The suppliers will have to adhere to the delivery schedule. Delayed supplies beyond 30 days from the date of supply order as the case may be will be subject to demurrage @0.5% per weeks or part thereof of the contract price subject to maximum of 10% of contract price, if the Price is below 2 lakh, else it will be @ 0.5% up to as maximum of 5%.
14. Warranty/Guarantee should be quoted as per the Technical specifications. (2 Years)
15. The successful bidder will submit P.B.G in the shape of FDR/TDR only in favor of **Dean, Shri Vasantrao Naik Government Medical College, Yavatmal @ 3% of the total value of the order** which will be valid for 30 months in case the warranty is for 24 months and 66 months in case the warranty is for five years. The same shall be further extended for 5 more year till the time AMC/CMC inforce.

16. The Manufacturer/Distributor/Agent should furnish along with the Price Bid a certificate that the rates quoted by the firm are the same as are being charged from other Government or Non-Government Departments/Institutions for similar items.


17. **Compliance Chart :-** (If quoting for more than one model then separate formats should be enclosed for different models in the same envelope.) Make & Model No of the quoted Item should be mentioned on the top Part of the compliance Chart.

Equipment Name -			Make & Model:	
Sr No	Tender /Quotation specifications as asked in the tender form	Equivalent specifications quoted by the tendered with the name of the manufacturer	Whether the tender/ Quotation quoted by the supplier is as per specifications asked for Indicate against each item clearly YES/NO	If the tender quoted is not as per specifications then variations/deviations should be clearly indicated against each item asked for.
1	2	3	4	5

18. The firm should submit undertaking on letter head stating that firm & its personnel do not have any direct or indirect interests with purchasing institute.
19. The Envelop No. 2 (Commercial Bid) will be opened of only those tenderers who have submitted all the documents as per tender terms and conditions in the envelop no. 01 and are technically qualified. The demonstration of tenderer who had quoted lowest price will be taken. If two tenderers had quoted same cost of equipment, then the demonstration of both the tenderer will be taken.
20. **Complaint/s :-** Any complaint / representation regarding tender/quotation will be entertained only after depositing of **Rs. 10,000/-** in form of **Demand Draft** in the name of Dean Shri V N Govt Medical College Yavatmal. Subsequently necessary action will be taken by the Dean and decision of Dean will be binding upon the complainant. If the complaint turns out to be false or invalid the amount will be forfeited. The Amount shall be refunded if after scrutiny the complaint is found to be true. No further complaint/ representation from the same complainant for the same tender/quotation will be entertained.

Note:- The Technical Specification, quoted price & availability of service Centre near Yavatmal will be considered while approving the prices.

21. I) The Instrument/Equipment Shall be of ISI/ISO mark or equivalent.
II) It shall necessary to provide OEM (original Equipment Manufacturer)
III) Supplier to provide user of maintenance AMC/CMC additional services.



DEAN


Shri Vasantrao Naik Government Medical
College, Yavatmal

IMPORTANT NOTE:- Every instrument & Equipment Quotation submit separate Envelop otherwise quotation rejected.

IFICATIONS FOR SINGLE DOOR REFRIGERATOR


- Refrigerator should be Single door with capacity of 400 liters
1. Frost free
 3. Climate class must be T Class, can work upto temp 45⁰ C.
 4. External dimension should be (WDH in mm) 600*680*1985.
 5. Internal dimension should be (WDH in mm) 518*475*1495
 6. Outer and Inner cabinet material should be of Metal type. No. shelves 5
 7. Door frame material should be of Plastic type with external handle& door lock
 8. Cooling fan motor Watt 18w, Condenser fan motor Watt 25w & Condenser type should be wire on tube
 9. Condenser coil type of material should be of steel& material for Evaporator coil Aluminium
 10. Material for Internal discharge& suction tube- COPPER with discharge & suction tube diameter 0.236 (inch)
 11. Power cord length should be of 2.5 meter.
 12. Compressor capacity should be with 215 watts
 13. Refrigerator type R134a with facility of Thermostat.
 14. Noise level (dBA) upto 43, Provision for Inner cabinet lamp -1 with power of 12w
 15. Input power watt 270 W and current 1.6Amps.
 16. Energy consumption 2.6 Kwh/24Hr with electricity safety approval
 17. WARRANTY L ONE year


Associate Prof. of Biochemistry
Shri. V. N. Govt. Medical College
KAVATMAL
वैद्यकीय विभाग
श्री. व. न. शासकीय वैद्यकीय महाविद्यालय
कावतमाल



प्राध्यापक व विभाग प्रमुख
जीवरसायनशास्त्र विभाग
श्री. व. न. शासकीय वैद्यकीय महाविद्यालय
कावतमाल

SPECIFICATIONS FOR MEDICAL CLINICAL CENTRIFUGE, BRUSHLESS

1. Centrifuge with provision of 48 holes of 5ml and 7ml vacutainers
- 2.
3. Table top, Low noise, Very stable, No vibration
4. Heavy duty C,R,C construction finished with powder coating
5. Fitted with Brushless motor
6. Heavy duty lid with hydraulic hinges and lid lock.
7. Swing out type Rotor Head
8. Centrifuge should be with stainless steel bowl of 304 quality
9. Facility for adopters of stainless tube for using 12x100ml tube of borosil
10. Required speed approx. 3000 rpm for serum separation
11. Provision of LCD Display of r.p.m, set time and run time.
12. Two lines of 20 character LCD display of RPM, Set Time and Run time
13. WARRANTY : 5 years on motor and 1 year on rest of items including electronic mother board



Prof. of Biochemistry
Shri. V. N. Govt. Medical College
YAVATMAL



जी.व.ना.शास्त्रीय वैद्यकीय महाविद्यालय
अवतमाल

जी.व.ना.शास्त्रीय वैद्यकीय महाविद्यालय
अवतमाल